

日本医師会臨床検査精度管理調査 2012

参加項目修正点 97.3	Specimen 1		Specimen 2		Specimen 3	
評価項目修正点 97.3	Result	SDI	Result	SDI	Result	SDI
Chemistry						
AST	23	-0.2	88	0.1	116	-0.4
ALT	24	0.5	84	1.2	110	0.8
ALP	104	-0.8	504	-0.5	678	-1.4
LD	145	0.0	393	0.2	495	-0.2
γ-GT	30	-0.2	152	1.0	204	0.8
CHE	130	0.2	349	0.2	440	-0.3
AMY	60	-0.6	223	-0.1	293	-0.3
CK	80	0.8	260	0.8	336	0.6
TP	7.4	0.1	4.7	0.0		
ALB	4.2	-0.5	2.2	-0.6		
GLU	89	0.0	201	0.2	250	0.2
BUN	10.2	0.2	40.8	0.4	54	0.3
CRE	0.81	0.4	3.1	1.1	4.07	1.0
UA	3.5	0.0	8.6	0.0	10.8	-0.2
T-Bil	1	0.0	4.2	0.9	5.6	1.0
T-CHO	214	0.0	117	-0.4	167	-0.1
HDL-C	59	1.1	35	0.7	50	0.8
LDL-C	131	-0.2	71	-0.7	103	-0.2
TG	176	0.4	85	0.0	122	0.3
IP	3	0.3	6.6	0.6	8.2	0.6
Ca	8	0.1	11.8	0.1	13.4	0.0
Fe	64	1.2	173	1.5	218	1.0
HbA1c	6	0.1	10.1	0.3		
Neoplasm marker & hormone						
Insulin	12.1	-0.6	53.4	0.0	97.5	-0.3
TSH	2.8	-0.5	8.6	-0.2	14.4	-0.1
FT4	1.98	0.9	2.67	1.4	3.4	1.7
PSA	2.8	0.0	16.9	0.3	30.5	-0.1
AFP	8	0.0	65	-0.2	118	-0.6
CEA	3.4	0.1	16.2	-0.1	29.1	0.0
CA19-9	24	0.2	66	0.0	110	-0.2
CA-125	24	0.2	60	1.2	88	0.4
Plasma protein & others						
CRP	0.24	0.1	1.75	0.6		
Ferritin	39	-1.2	195	-0.8	338.8	-1.5
RF	3	-	31.3	0.4	62	0.7
TPLA	positive	positive	negative	negative		
HBs-Ag	positive	positive	negative	negative		
HCV-Ab	positive	positive	negative	negative		
Hematology						
WBC	117	0.0	30	-0.3		
RBC	414	-0.1	484	-0.3		
Hgb	11.4	0.3	14.5	0.6		
Hct	33.1	0.2	41	0.1		
Plt	14.4	-1.2	30.9	-1.2		
Coagulation						
PT ratio	2.53	-				
PT INR	2.58	1.5				
APTT	30.9	1.7	40.4	0.2		
Fibrinogen	221	-0.1	253	0.0		
Clinical Microscopy						
Protein, urine	陰性	陰性	100mg/dL	100mg/dL		
Glucose, urine	陰性	陰性	30mg/dL	30mg/dL		
Blood/Hgb,urine	陰性	陰性	2+以上	2+以上		