

日本医師会臨床検査精度管理調査 2009

参加項目修正点	90.6	Specimen 1		Specimen 2		Specimen 3	
評価項目修正点	95.4	Result	SDI	Result	SDI	Result	SDI
Chemistry							
AST		27	1.8	79	0.4	122	-0.2
ALT		23	0.6	75	1.4	116	0.8
ALP		112	0.1	432	0.6	693	-0.4
LD		137	-0.2	356	1.1	518	0.4
-GT		27	-2.2	127	-0.9	206	-1.0
CHE		128	0.5	315	1.1	457	0.4
AMY		60	0.1	193	0.0	300	-0.1
CK		82	0.3	228	0.4	345	-0.1
GLU		86	0.1	186	1.2	259	-0.2
BUN		10.1	0.3	35.8	0.2	55.9	-0.6
CRE		0.8	0.3	2.76	1.1	4.26	0.6
UA		3.4	-0.5	7.6	-0.6	10.9	-1.0
T-Bil		0.6	-0.9	2.4	0.3	3.7	-0.5
T-CHO		190	-0.5	117	-0.8	166	-0.8
HDL-C		46	-0.3	34	-0.7	49	-0.5
LDL-C		92	-0.1	72	-0.6	102	-0.9
TG		158	0.6	85	0.4	121	0.3
IP		3.3	-1.1	6.2	-0.9	8.5	-1.0
Ca		8.2	0.3	11.2	0.3	13.5	0.0
Fe		63	0.1	170	0.4	122	-0.2
HbA1c		5.8	-0.4	9.5	-0.9		
Neoplasm marker & hormone							
TSH		3.9	1.0	8	2.6	23.1	2.1
FT4		1.71	-0.8	2.5	-0.3	5.97	0.0
PSA		2.1	-1.3	5.2	-0.7	17.7	-0.2
AFP		9	-0.5	34	1.1	119	-0.7
CEA		3.7	-0.4	8.5	-0.6	28.4	-0.4
CA19-9		27	0.2	59	0.0	206	0.0
CA-125		19	-0.5	34	-0.4	93	-0.7
Plasma protein & others							
IgG		1325	-0.8	1825	-0.7		
IgA		234	-1.1	316	-2.1		
IgM		124	-0.1	173	-0.5		
CRP		0.6	1.1	4.76	1.1		
HBs-Ag		negative	negative	positive	positive		
HCV-Ab		positive	positive	negative	negative		
Hematology							
WBC		49	0.5	120	-0.1		
RBC		358	-0.3	426	0.1		
Hgb		10.8	-0.2	11.9	0.2		
Het		31.2	-0.6	34.9	0.4		
Plt		12	-0.3	15.1	-0.5		
Coagulation							
PT ratio		2.43					
PT INR		2.56					
APTT		35.3		61.6			
Fibrinogen		271		305			