

日本医師会臨床検査精度管理調査2014

参加項目修正点 98.8	Specimen 1		Specimen 2		Specimen 3	
評価項目修正点 98.8	Result	SDI	Result	SDI	Result	SDI
Chemistry						
AST	23	-0.3	55	-0.1	115	-0.3
ALT	23	-0.6	53	-0.6	110	-0.1
ALP	177	-0.2	364	-1.0	757	-0.9
LD	138	-0.1	258	-0.8	479	-867.0
γ-GT	31	0.6	88	-0.2	195	-0.3
CHE	142	0.3	248	-0.1	446	-0.3
AMY	61	-0.2	142	-0.6	297	-0.3
CK	78	-0.3	159	-0.7	312	-0.3
TP	7.6	0.1	4.9	0.3		
ALB	4.3	-0.8	2.3	0.0		
GLU	91	0.1	145	-0.2	249	-0.1
BUN	10.1	0.3	25.1	-0.1	54.4	0.3
CRE	0.76	0.1	1.86	0.3	3.88	-0.2
UA	3.5	0.0	6	-0.3	10.8	-0.3
T-Bil	0.9	0.0	2.4	0.9	5	0.5
T-CHO	225	0.7	125	0.8	176	0.5
HDL-C	58	0.4	35	0.8	49	0.3
LDL-C	140	0.9	78	0.9	112	1.1
TG	157	0.2	92	0.1	132	0.3
IP	3.2	0.3	4.8	-0.3	8	0.0
Ca	8	-0.3	9.8	-0.2	13.3	0.0
Fe	73	1.1	133	0.7	246	0.5
HbA1c	6	0.3	10.1	0.0		
Neoplasm marker & hormone						
Insulin	14.4	-0.7	57	-0.8	93.9	-0.5
TSH	3.5	0.3	11.5	-257.0	18.4	0.2
FT4	1.78	0.1	2.69	0.1	3.5	-0.2
PSA	2.7	0.5	17.7	0.3	30.5	0.6
AFP	6	0.0	54	-0.5	99	0.7
CEA	5.1	0.2	24.5	-0.1	41.3	0.4
CA19-9	19	0.4	128	0.3	227	0.6
CA-125	15	0.0	45	-0.2	71	0.3
Plasma protein & others						
CRP	0.29	-0.7	1.79	-1.2		
Ferritin	58.4	-1.6	266.4	-0.8	425.2	-1.2
RF	5	-	46.9	1.6	85	1.5
TPLA	positive	positive	negative	negative		
HBs-Ag	positive	positive	negative	negative		
HCV-Ab	positive	positive	negative	negative		
Hematology						
WBC	122	0.5	30	0.3		
RBC	347	0.1	483	0.1		
Hgb	8.8	-0.2	14	-0.2		
Hct	26.1	0.0	40	0.2		
Plt	48.1	-0.9	31	-0.9		
Coagulation						
PT ratio	2.58	-				
PT INR	2.56	-0.3				
APTT	29.3	0.9	39.9	0.2		
Fibrinogen	237	0.6	275	0.5		
Clinical Microscopy						
Protein, urine	陰性	陰性				
Glucose, urine	陰性	陰性				
Blood/Hgb,urine	陰性	陰性				